



Uniphos Envirotronic Pvt.Ltd. - 100% EOU

C.I.N. U29297GJ2007PTC052025

36 (1), Jalaram Mandir Road,
Village - Nahuli, Via-Vapi,
Taluka - Umbergaon, Dist. Valsad.
Pin Code - 396 108. Gujarat.
Tel.: +91-260-2730156
+91-9909994042, 7574839945
E-mail : uepl@uniphos-envirotronic.com
Website : www.uniphos-she.com

A UPL Limited Group Company

Date 04/09/2017

TO WHOMSOEVER IT MAY CONCERN

This is to certify that the below mentioned KWIKDRAW/UNIPHOS Gas Detector Tubes are manufactured at Uniphos Envirotronic Pvt. Ltd (100% EOU), an ISO 9001: 2008 certified facility. These products have been manufactured according to UNIPHOS's Quality Assurance Plan. It is further certified that, these tubes comply with the requirements of the European Pharmacopoeia 6.0.

Part No.	Description	MOC and Indicating material	Minimum value indicated	Relative standard deviation
D5086814	Carbon dioxide detector tube CO2 -100	Sealed glass tube containing adsorbent fillers and suitable support for hydrazine and crystal violet indicator.	100 ppm	±15 %
D5085803	Sulphur dioxide detector tube SO2- 1	Sealed glass tube containing adsorbent fillers and suitable support for iodine and starch indicator.	0.5 ppm	±15 %
10040887	Oil Detector tube Oil-HP synthetic	Sealed glass tube containing adsorbent fillers and suitable support for sulphuric acid indicator.	0.1 mg/m ³	±30 %
D5085818	Nitrogen monoxide and Nitrogen dioxide detector tube Nitr-0.5	Sealed glass tube containing adsorbent fillers and suitable support for an oxidising layer (Cr(VI) salt) and the diphenylbenzidine indicator.	0.5 ppm	±15 %
D5085836	Carbon monoxide detector tube CO-5	Sealed glass tube containing adsorbent fillers and suitable support for di-iodine pentoxide, selenium dioxide and fuming sulphuric acid indicator.	5 ppm	±15 %
D5085826	Hydrogen sulphide detector tube H2S-1	Sealed glass tube containing adsorbent fillers and suitable support for an appropriated lead salt indicator.	1 ppm	±10 %
CF000712	Water vapour detector tube SWA-4L	Sealed glass tube containing adsorbent fillers and suitable support for the magnesium perchlorate indicator.	42.79 ppm at 20 °C	±20 %




Uniphos Envirotronic Pvt.Ltd. - 100% EOU

C.I.N. U29297GJ2007PTC052025

36 (1), Jalaram Mandir Road,
Village - Nahuli, Via-Vapi,
Taluka - Umbergaon, Dist. Valsad.
Pin Code - 396 108. Gujarat.
Tel.: +91-260-2730156
+91-9909994042, 7574839945
E-mail : uepl@uniphos-envirotronic.com
Website : www.uniphos-she.com

A UPL Limited Group Company

For Uniphos Envirotronic Pvt. Ltd. (100% EOU)



DR. Ajayakumar K.S.